

Certificate

on the antimicrobial properties of materials
The product

MEDI grip

of the company

HEINRICH KIPP WERK GmbH & Co. KG
Heubergstraße 2, 72172 Sulz am Neckar, Deutschland

has been successfully tested against
Staphylococcus epidermidis DSM 18857
Pseudomonas aeruginosa DSM 939 / ATCC 15442

as **antimicrobial**
(≥ 99.9% germ reduction)

according to ASTM E 2180 (modified)

For details, see Test Reports No. 2202.2 / 150922-683-2180-02; sample: 6832109154083;
2339.3 / 160301-10195-2180-03; sample: 101952902160021

The bearer of this certificate is entitled to affix the following seal to
parts or surfaces made from the tested material



Nuremberg, Germany, September 10th, 2020


Managing Director


Laboratory Director

QualityLabs BT GmbH, Neumeyerstraße 46a, 90411 Nuremberg, Germany

CERTIFIED according to ASTM E 2180